

Date/Time: 10/24/2018 5:59:14 PM  
 Username: ADMIN

# UDP Report

Instrument #: CT0156(1)  
 Version #: V4.70 A

## 400008CMAX S4

Test Name	Version	Extraction Type	Master Mix Format					Ct Calculation	Lysis Heat Time
Notes	Channel Settings	475/520	530/565	585/630	630/665	680/715	Lysis Temperature	Sample Tip Height	
							Sample Volume	Wash Volume	
							Neutralization Volume	DNase Heat Time	
							Liquid Level Sensing Enabled		
							400008CMAX S4	1	ExK DNA-2 - 4-Snap
--		Alias	--	VZV	HSV2	HSV1	SPC	62	1600
		Gain	--	60	60	60	60	937.5	500
		Threshold	--	200	200	200	100	--	--
		Excitation Channel 475/520	--	0	0	0	0		
		Excitation Channel 530/565	0	--	0	0	0		
		Excitation Channel 585/630	0	0	--	5	0		
		Excitation Channel 630/665	0	0	5	--	0		
		Excitation Channel 680/715	0	0	0	3	--		

Enabled

**Step Name:** New Step  
**Profile Type:** Hold **Cycles:** 1

Type	Time (s)	Temp (°C)	Detect
--	300	99	No

**Step Name:** New Step  
**Profile Type:** 2 - Temperature **Cycles:** 45

Type	Time (s)	Temp (°C)	Detect
--	13.7	99	No
--	46.9	64	Yes

	HSV1			
HSV1 - PCR	✓	✓	✗	✗
SPC - PCR	✓	✗	✓	✗
Results	➕ POS	➕ POS	➖ NEG	⚠ UNR

	HSV2			
HSV2 - PCR	✓	✓	✗	✗
SPC - PCR	✓	✗	✓	✗
Results	➕ POS	➕ POS	➖ NEG	⚠ UNR

	VZV			
VZV - PCR	✓	✓	✗	✗
SPC - PCR	✓	✗	✓	✗
Results	➕ POS	➕ POS	➖ NEG	⚠ UNR



Performed By: \_\_\_\_\_

Date: \_\_\_\_\_

Printed: 2018-10-24 17:59:15

Reviewed By: \_\_\_\_\_

Date: \_\_\_\_\_

Page: 1/1